



## CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.  
4356 Communications Drive  
Norcross, Georgia 30093-2901 U.S.A.  
ISO-9001 Certified

### VERIFICATION

**Model Number:** ASAP 2050

**Type of Equipment:** Surface Area Analyzer

**Standards of Conformity:** EN 61326: 1997 +A1: 1998  
IEC 61000-4-2 Electronic Discharge  
IEC 61000-4-3 RF Electromagnetic Field  
IEC 61000-4-4 Electrical Fast Transient/Burst  
IEC 61000-4-5 Surge  
IEC 61000-4-6 RF Common Mode  
IEC 61000-4-8 Surge  
IEC 61000-4-11 Voltage Dips, Interruptions, and Variations  
EN 61010-1 Part 1: General Safety Requirements

**Other Tests:** CISPR 11 Radiated Emissions  
CISPR 11 Conducted Emissions  
EN 61000-3-2 Harmonic Current Emissions  
EN 61000-3-3 Voltage Fluctuations

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

Stefan Koch  
Signature

2-24-2009  
Date

Stefan Koch  
Print name

Director of Engineering  
Position