

**characteristic mass for peak absorption (in electrothermal atomization),  $m_p$**

That mass of analyte which produces a defined peak absorbance of 0.0044 (or 1% absorbance).

1992, 64, 258