

**precision (of a balance)**

The *standard deviation* of the instrument for the stated load (e.g.  $l_{20}$  for a 20 g load). A statement of the procedure, conditions, and experience of the observer should be included.

O.B. 36