

He

AccuPyc 1330 V2.04N
Serial Number: 3845
Density and Volume Report

Sample ID: 06-0895 X33756-65
Sample Weight: 3.6421 g
Temperature: 23.9 C
Number of Purges: 30
Cell Volume: 12.0385 cm3
Chamber Insert: None

Started: 07/03/06 14:40:07
Completed: 07/03/06 15:10:35

Equilibration Rate: 0.0050 psig/min
Expansion Volume: 9.2666 cm3

| Run# | Volume cm3 | Deviation cm3 | Density g/cm3 | Deviation g/cm3 | Elapsed Time (h:m:s) |
|------|---------------|------------------|------------------|--------------------|-------------------------|
| 1 | 3.2422 | 0.0006 | 1.1233 | -0.0002 | 0:21:28 |
| 2 | 3.2416 | -0.0000 | 1.1236 | 0.0000 | 0:23:43 |
| 3 | 3.2413 | -0.0003 | 1.1236 | 0.0001 | 0:25:57 |
| 4 | 3.2421 | 0.0005 | 1.1234 | -0.0002 | 0:28:05 |
| 5 | 3.2408 | -0.0008 | 1.1238 | 0.0003 | 0:30:20 |

Average Volume: 3.2416 cm3
Average Density: 1.1236 g/cm3

Standard Deviation: 0.0006 cm3
Standard Deviation: 0.0002 g/cm3