



Particle Size Surface Area Pore Size Distribution Density

CE DECLARATION OF CONFORMITY

Micromeritics Instrument Corporation, Inc.
One Micromeritics Drive
Norcross, Georgia 30093-1877 U.S.A.
ISO-9001 Certified

VERIFICATION

Model Number: ASAP 2020

Type of Equipment: Surface Area Analyzer

Standards of Conformity: **EN 61326: 1997 +A1: 1998**
IEC 61000-4-2 Electronic Discharge
IEC 61000-4-3 RF Electromagnetic Field
IEC 61000-4-4 Electrical Fast Transient/Burst
IEC 61000-4-5 Surge
IEC 61000-4-6 RF Common Mode
IEC 61000-4-8 Surge
IEC 61000-4-11 Voltage Dips, Interruptions, and Variations
EN 61010-1 Part 1: General Safety Requirements

Other Tests: CISPR 11 Radiated Emissions
CISPR 11 Conducted Emissions
EN 61000-3-2 Harmonic Current Emissions
EN 61000-3-3 Voltage Fluctuations

Micromeritics Instrument Corporation hereby declares that the instrument(s) specified above conform(s) to the Directive(s) and Standard(s), when installed and operated in accordance with the specifications set forth by Micromeritics. The original of this document is kept at Micromeritics with copies of the relevant test data and certificates, which constitute the required technical file for self declaration.

G. R. Killip
Signature

28th Feb 2003
Date

GRAHAM KILLIP
Print name

DIRECTOR OF ENGINEERING
Position