



## UK DECLARATION OF CONFORMITY

This declaration of conformity is issued under the sole responsibility of the manufacturer:

**Micromeritics Instrument Corporation**  
4356 Communications Drive  
Norcross, GA 30093, USA

Hereby declares that the product:

**FlowPrep Sample Preparation System**  
**Model FlowPrep (060)**

is in conformity with the following UK legislation:

**Electrical Equipment (Safety) Regulations 2016**  
**Electromagnetic Compatibility Regulations 2016**  
**Restriction of the Use of Certain Hazardous Substances in E&E Equipment Regulations 2012**

and that the equipment is in conformity with the following designated and other appropriate standards;

**Electrical Equipment (Safety) Regulations 2016**

**EN 61010-1:2010 + A1:2019** - *Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements.*

**Electromagnetic Compatibility Regulations 2016**

**EN 61326-1:2013** - *Electrical equipment for measurement, control and laboratory use — EMC requirements — Part 1: General requirements*

**IEC 61000-3-2:2014** - *Part 3-2: Limits — Limits for harmonic current emissions (equipment input current  $\leq 16$  A per phase)*

**IEC 61000-3-3:2013** - *Part 3-3: Limits — Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current  $\leq 16$  A per phase and not subject to conditional connection*

**Restriction of the Use of Certain Hazardous Substances in E&E Equipment Regulations 2012**

**EN 63000: 2018** - *Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances*

Name: John McCaffrey, Ph.D.

Title: Vice President, R & D

Signature: 

Date of issue: 03/30/2021

Location: Norcross, GA USA