

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

## CB TEST CERTIFICATE

## Product

Laboratory Equipment Volume and Density Analyzer

## Name and address of the applicant

Micromeritics Instrument Corporation  
4356 Communications Drive  
Norcross, GA, 30093  
United States of America

## Name and address of the manufacturer

Micromeritics Instrument Corporation  
4356 Communications Drive  
Norcross, GA, 30093  
United States of America

## Name and address of the factory

*Note: When more than one factory, please report on page 2*

Micromeritics Instrument Corporation  
4356 Communications Drive  
Norcross, GA, 30093  
United States of America

Additional information on page 2

## Ratings and principal characteristics

Marked On Device: 24VDC, 120W; Powered By An External Power Supply  
Rated 100-240Vac, 50-60Hz, 1.6A; Output 24Vdc, 5A, 120W

## Trademark / Brand (if any)



micromeritics

## Customer's Testing Facility (CTF) Stage used

## Model / Type Ref.

GeoPyc 1365

## Additional information (if necessary may also be reported on page 2)

Additional information on page 2

## A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010/AMD1:2016, IEC 61010-2-081:2019

Comments:

Refer to the Test Report for the Summary of Compliance with National Deviations

## As shown in the Test Report Ref. No. which forms part of this Certificate

105317335ATL-001C, 105317335ATL-001D

This CB Test Certificate is issued by the National Certification Body

Intertek Testing Services NA, Inc.  
545 E. Algonquin Road  
Arlington Heights  
IL 60005  
United States of America

Date: 2023-03-22

Signature: John Quigley